

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0175980	(X3) Date Survey Completed 10/10/2019
Name of Provider or Supplier Medi-Help Pc Laboratory	Street Address, City, State 1691 Washington Rd, Mount Lebanon, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2087	<p>ROUTINE CHEMISTRY CFR(s): 493.841(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on interview of the Laboratory Director and review of the proficiency test records, at the time of the survey, the laboratory failed to attain a score of at least 80, for 1 of 20 analytes in the American Proficiency Institute 2019 2ND testing event. Findings: 1. Chloride proficiency test score, was 20% for the American Proficiency Institute 2019 2ND testing event. 2. During the survey (15:30 10/10/2019), the the Laboratory Director confirmed the above finding.</p>