

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0179820	(X3) Date Survey Completed 08/15/2019
Name of Provider or Supplier Cornerstone Care Ped Ctr Of Washington	Street Address, City, State 351 West Beau St, Ste 200, Washington, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6054	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least annually, after the first year.</p> <p>This STANDARD is not met as evidenced by: Based on review of the competency records and personnel interview of the Testing Personnel #1 and the Practice Manager, on (08/15/2019), the Technical Consultant failed to document Testing Personnel competency at least annually for one of two years, (04/26/2018 through 08/15/2019). Findings include: 1. The Testing Personnel competency was last documented 04/25/2018. 2. 6 of 6 Testing Personnel competency assessments were not found for 2019. 3. During the survey (14:10 08/15/2019) the Practice Manager confirmed the above findings.</p>