

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  39D0185151	<b>(X3) Date Survey Completed</b>  07/12/2023
<b>Name of Provider or Supplier</b>  Ph Dubois, Clearfield Campus	<b>Street Address, City, State</b>  809 Turnpike Avenue, Clearfield, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2006</b>	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)</p> <p>The laboratory must examine or test, as applicable, the proficiency testing samples it receives from the proficiency testing program in the same manner as it tests patient specimens. This testing must be conducted in conformance with paragraph (b)(4) of this section. If the laboratory's patient specimen testing procedures would normally require reflex, distributive, or confirmatory testing at another laboratory, the laboratory should test the proficiency testing sample as it would a patient specimen up until the point it would refer a patient specimen to a second laboratory for any form of further testing.</p> <p>This STANDARD is not met as evidenced by: Based on review of the American Proficiency Institute (API) proficiency testing (PT) records and interview with technical supervisor #4 (TS), the laboratory failed to examine 1 of 6 API PT hematology testing events in the same manner as patient specimens in 2021, 2022, and 2023. Findings include: 1. On the day of survey, 07/11/2023 at 10:06 am, review of the patient test reports and the API PT records revealed that the laboratory failed to examine the following PT samples for 1 of 6 API PT hematology testing events in the same manner as patient specimens in 2021, 2022, and 2023: -2023 API Heme/Coag 1st Event - Body Fluid Crystals - Semen Analysis - Cell Identification 2. During an interview, 07/11/2023 at 10:30 am, TS #4 stated, "PT samples for cell identification, body fluid, and semen analysis were analyzed by two different testing personnel." 3. The laboratory could not provide documentation stating the tests listed above where included in the laboratory's patient repeat criteria. 4. TS #4 confirmed the findings above on 07/12/2023 around 12:00 pm.</p>
<b>D3009</b>	<p>FACILITIES CFR(s): 493.1101(c)</p>

The laboratory must be in compliance with applicable Federal, State, and local laboratory requirements.

This STANDARD is not met as evidenced by:

Based on record review and interview with the technical supervisor (TS) #2, the laboratory failed to ensure a general supervisor (GS) who met the state requirement was on the laboratory premises during all hours in which tests were performed from 10/07/2020 through the day of survey. Findings include 1) The PA regulations (5.23(b) (1) states: "A general supervisor who meets all the requirements of subsection (a)(1), (2) or (3) and is on the laboratory premises during all normal scheduled working hours in which tests are being performed." 2) On the day of the survey, 07/11/2023 at 8:30 am review of Laboratory Personnel Report (Pennsylvania State) and personnel credentials, revealed GS #3 does not meet the education and experience requirements to supervise testing in the laboratory. 3) During an interview with TS #2, it was discovered the laboratory does not have a qualified supervisor onsite for every hour of patient testing according to chapter 5 section 5.23 of the Pennsylvania State regulations for clinical laboratories. 4) On 07/12/2023 TS#2 confirmed the above findings around 12:00 pm

**D3013**

**FACILITIES**

CFR(s): 493.1101(e)

Records and, as applicable, slides, blocks, and tissues must be maintained and stored under conditions that ensure proper preservation.

This STANDARD is not met as evidenced by:

Based on observation of the laboratory and interview with technical supervisor #2 (TS), the laboratory failed to record the room temperature for paraffin block storage to ensure proper preservation from 10/07/2020 to the day of the survey. Findings include: 1. On the day of the survey, 07/12/2023 at 10:30 AM, during the laboratory tour it was observed that the paraffin blocks were not stored in a temperature monitored area. 2. The laboratory could not provide the room temperature records for paraffin blocks stored within two sections of the laboratory hallway from 10/07/2020 to the day of the survey. 3. TS #2 confirmed the finding above on 07/12/2023 around 12:00 PM.

**D5403**

**PROCEDURE MANUAL**

CFR(s): 493.1251(b)

The procedure manual must include the following when applicable to the test procedure: (1) Requirements for patient preparation; specimen collection, labeling, storage, preservation, transportation, processing, and referral; and criteria for specimen acceptability and rejection as described in 493.1242. (2) Microscopic examination, including the detection of inadequately prepared slides. (3) Step-by-step performance of the procedure, including test calculations and interpretation of results. (4) Preparation of slides, solutions, calibrators, controls, reagents, stains, and other materials used in testing. (5) Calibration and calibration verification procedures. (6) The reportable range for test results for the test system as established or verified in 493.1253. (7) Control procedures. (8) Corrective action to take when calibration or control results fail to meet the laboratory's criteria for acceptability. (9) Limitations in the test methodology, including interfering substances. (10) Reference intervals

(normal values). (11) Imminently life-threatening test results, or panic or alert values. (12) Pertinent literature references. (13) The laboratory's system for entering results in the patient record and reporting patient results including, when appropriate, the protocol for reporting imminently life threatening results, or panic, or alert values. (14) Description of the course of action to take if a test system becomes inoperable.

This STANDARD is not met as evidenced by:

A. Based on review of the laboratory's procedures and interview with technical supervisor #2 (TS), the laboratory failed to include, step-by-step performance procedures for immunohematology testing and for general laboratory operations from 10/07/2020 to 07/12/2023. Findings include: 1. On the days of the survey, 7/11/2023 and 7/12/2023, the laboratory could not provide the following procedures: - Immunohematology: 1. ABO Discrepancy Procedure -General Laboratory Operations: 1. New Method and Test Validation Procedure 2. Individualized Quality Control Plan (IQCP) Procedure 2.The TS#2 confirmed the finding above on 07/12/2023 at 12:00 pm. B. Based on the review of the laboratory's procedures and interview with technical supervisor #2 (TS), the laboratory failed to ensure the procedures included frequency instructions for quality control processes and the laboratory's system for entering results in the patient records in hematology and microbiology from 10/07/2020 to 7/12/2023. Findings include: 1. On the days of the survey, 7/11/2023 and 7/12/2023, the laboratory failed to include the frequency of the quality control processes in the procedure for spot indole in microbiology. 2.The laboratory failed to include the step-by-step process for reporting patients results in hematology and microbiology for the following procedures: -Cerebral Spinal Fluid (CSF) -Body Fluids 3.TS#2 confirmed the above findings on 7/12/2023 at 12:00 PM.

**D5421**

**ESTABLISHMENT AND VERIFICATION OF PERFORMANCE**  
CFR(s): 493.1253(b)(1)

Each laboratory that introduces an unmodified, FDA-cleared or approved test system must do the following before reporting patient test results: (1)(i) Demonstrate that it can obtain performance specifications comparable to those established by the manufacturer for the following performance characteristics: (1)(i)(A) Accuracy. (1)(i)(B) Precision. (1)(i)(C) Reportable range of test results for the test system. (1)(ii) Verify that the manufacturer's reference intervals (normal values) are appropriate for the laboratory's patient population.

This STANDARD is not met as evidenced by:

Based on a review of the Sysmex XN-2000 analyzer validation records and an interview with technical supervisor #2 (TS), the laboratory failed to establish criteria for acceptable performance specifications and perform a reference range/normal value study for 13 of 13 analytes performed on the Sysmex XN-2000 hematology analyzer from 9/29/2021 to the date of the survey. Findings Include: 1. On the day of the survey, 07/11/2023 at 2:30 PM, a review of the Sysmex XN-2000 analyzer validation records revealed the validation performed on 9/21/2021 did not include the laboratory's acceptable criteria for performance specifications for precision, accuracy, and reportable ranges for the following analytes: - White Blood Count (WBC) - Red Blood Count (RBC) - Hemoglobin (HGB) - Hematocrit (HCT) - Mean Corpuscular Volume (MCV) - Mean Corpuscular Hemoglobin (MCH) - Mean Corpuscular Hemoglobin Concentration (MCHC) - Mean Platelet Volume (MPV) - Red Blood Cell Distribution Weight (RDW) - Platelet (PLT) - Neutrophils (NEU) - Lymphocytes

(LYM) - Monocytes (MON) 2. The laboratory could not provide documentation of a reference range/normal value study appropriate for the laboratory's patient population. 3. TS#2 confirmed the findings above on 07/12/2023 around 12:00 PM.

**D5775**

**COMPARISON OF TEST RESULTS**

CFR(s): 493.1281(a)(c)

(a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites. (c) The laboratory must document all test result comparison activities.

This STANDARD is not met as evidenced by:

Based on lack of documentation and interview with technical supervisor #2 (TS), the laboratory failed to have a system in place to evaluate and define the relationship between manual urine microscopic tests and automated urine microscopic tests (Arkray AU-4050) twice a year from 10/07/2020 to 07/12/2023. Findings Include: 1. On the day of survey, 07/11/2023 at 11:07am, the laboratory could not provide comparison studies between manual urine microscopic tests and the automated urine microscopic tests performed on the Arkray AU-4050 at least twice a year from 10/07/2020 to 07/12/2023. 2. TS#2 confirmed the finding above on 07/12/2023 at 12:00 pm.

**D5807**

**TEST REPORT**

CFR(s): 493.1291(d)

Pertinent "reference intervals" or "normal" values, as determined by the laboratory performing the tests, must be available to the authorized person who ordered the tests and, if applicable, the individual responsible for using the test results.

This STANDARD is not met as evidenced by:

Based on record review and interview with the technical supervisor (TS) #2, the laboratory failed to include reference intervals or normal values for chemistry and hematology from 10/07/2020 to 07/12/2023. Findings include: 1. On the day of survey 07/12/2023 at 08:30 am, review of the patient records revealed that the following records did not include reference intervals or normal values for the following analytes from 10/07/2020 to 07/12/2023: - alcohol levels (ETOH) - synovial crystals 2. The TS#2 confirmed the finding above on 07/12/2023 at 12:00 pm.

**D6093**

**LABORATORY DIRECTOR RESPONSIBILITIES**

CFR(s): 493.1445(e)(5)

The laboratory director must ensure that the quality control programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur.

This STANDARD is not met as evidenced by:

Based on review of quality control (QC) records and interview with the technical

supervisor (TS)#2, the laboratory director (LD) failed to ensure that the QC programs were established and maintained for in microbiology, hematology, chemistry, urinalysis, and Immunoematology from 10/07/2020 to the day of survey. Findings include: 1. On the days of survey 07/11/2023 and 07/12/2023, review of QC records in microbiology, hematology, chemistry, urinalysis, and Immunoematology revealed the following: a. Microbiology: - No negative control for Streptex serotyping - Individualized quality control plan (IQCP) is not separated by media. - IQCP for microscan is not separated by panel - Microbiology media QC records do not specify the biochemical reactions and suppression of organisms. - QIASTAT no QC records from 03/2021 to 10/2021 b. Hematology: - Testing personnel are not following body fluid QC procedures. c. Microbiology, Hematology, Chemistry, Urinalysis, and Immunoematology: - QC is not being rotated among all testing personnel who perform patient testing. 2. The TS #2 confirmed the findings above on 07/12/2023 at 12:00 pm

**D6111**

**TECHNICAL SUPERVISOR QUALIFICATIONS**  
CFR(s): 493.1449

(a) The technical supervisor must possess a current license issued by the State in which the laboratory is located, if such licensing is required; and (b) The laboratory may perform anatomic and clinical laboratory procedures and tests in all specialties and subspecialties of services except histocompatibility and clinical cytogenetics services provided the individual functioning as the technical supervisor-- (b)(1) Is a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (b)(2) Is certified in both anatomic and clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or Possesses qualifications that are equivalent to those required for such certification. (c) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of bacteriology, the individual functioning as the technical supervisor must-- (c)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (c)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (c)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (c)(2)(ii) Have at least one year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of bacteriology; or (c)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (c)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of bacteriology; or (c)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (c)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of bacteriology; or (c)(5)(i) Have earned a bachelor's degree in a chemical, physical, or biological science or medical technology from an accredited institution; and (c)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum

of 6 months experience in high complexity testing within the subspecialty of bacteriology. (d) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of mycobacteriology, the individual functioning as the technical supervisor must-- (d)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (d)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (d)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (d)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycobacteriology; or (d)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (d)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycobacteriology; or (d)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (d)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycobacteriology; or (d)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (d)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycobacteriology. (e) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of mycology, the individual functioning as the technical supervisor must-- (e)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (e)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (e)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (e)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycology; or (e)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (e)(3)(ii) Have at least 1 year of laboratory training or experience, or both in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycology; or (e)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (e)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycology; or (e)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (e)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the

specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of mycology. (f) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of parasitology, the individual functioning as the technical supervisor must-- (f)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (f)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (f)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (f)(2)(ii) Have at least one year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of parasitology; (f)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (f)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of parasitology; or (f)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (f)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of parasitology; or (f)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (f)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of parasitology. (g) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of virology, the individual functioning as the technical supervisor must-- (g)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (g)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (g)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (g)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of virology; or (g)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (g)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of virology; or (g)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (g)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing within the specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of virology; or (g)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (g)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing within the

specialty of microbiology with a minimum of 6 months experience in high complexity testing within the subspecialty of virology. (h) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of diagnostic immunology, the individual functioning as the technical supervisor must-- (h)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (h)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (h)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (h)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing for the specialty of diagnostic immunology; or (h)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (h)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of diagnostic immunology; or (h)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (h)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing for the specialty of diagnostic immunology; or (h)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (h)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing for the specialty of diagnostic immunology. (i) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of chemistry, the individual functioning as the technical supervisor must-- (i)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (i)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (i)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (i)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing for the specialty of chemistry; or (i)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (i)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of chemistry; or (i)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (i)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing for the specialty of chemistry; or (i)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (i)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing for the specialty of chemistry. (j) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of hematology, the individual functioning as the technical supervisor must-- (j)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (j)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (j)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (j)(2)(ii) Have at least

one year of laboratory training or experience, or both, in high complexity testing for the specialty of hematology (for example, physicians certified either in hematology or hematology and medical oncology by the American Board of Internal Medicine); or (j) (3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (j)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of hematology; or (j)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (j)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing for the specialty of hematology; or (j) (5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (j)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing for the specialty of hematology. (k)(1) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of cytology, the individual functioning as the technical supervisor must-- (k)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (k)(1)(ii) Meet one of the following requirements-- (k)(1)(ii)(A) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (k)(1)(ii) (B) Be certified by the American Society of Cytology to practice cytopathology or possess qualifications that are equivalent to those required for such certification; (l) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of histopathology, the individual functioning as the technical supervisor must-- (l)(1) Meet one of the following requirements: (l)(1)(i)(A) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (l)(1)(i)(B) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; (l)(1)(ii) An individual qualified under 493.1449(b) or paragraph (l)(1) of this section may delegate to an individual who is a resident in a training program leading to certification specified in paragraph (b) or (l)(1)(i)(B) of this section, the responsibility for examination and interpretation of histopathology specimens. (l)(2) For tests in dermatopathology, meet one of the following requirements: (l)(2)(i)(A) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located and-- (l) (2)(i)(B) Meet one of the following requirements: (l)(2)(i)(B)(1) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (l)(2)(i)(B)(2) Be certified in dermatopathology by the American Board of Dermatology and the American Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (l)(2)(i)(B) (3) Be certified in dermatology by the American Board of Dermatology or possess qualifications that are equivalent to those required for such certification; or (l)(2)(ii) An individual qualified under 493.1449(b) or paragraph (l)(2)(i) of this section may delegate to an individual who is a resident in a training program leading to certification specified in paragraphs (b) or (l)(2)(i)(B) of this section, the responsibility for examination and interpretation of dermatopathology specimens. (l) (3) For tests in ophthalmic pathology, meet one of the following requirements: (l)(3)(i) (A) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located and-- (l)(3)(i)(B) Must meet one of the following requirements: (l)(3)(i)(B)(1) Be certified in anatomic pathology

by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (l)(3)(i)(B)(2) Be certified by the American Board of Ophthalmology or possess qualifications that are equivalent to those required for such certification and have successfully completed at least 1 year of formal post-residency fellowship training in ophthalmic pathology; or (l)(3)(ii) An individual qualified under 493.1449(b) or paragraph (1)(3)(i) of this section may delegate to an individual who is a resident in a training program leading to certification specified in paragraphs (b) or (1)(3)(i)(B) of this section, the responsibility for examination and interpretation of ophthalmic specimens; or (m) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the subspecialty of oral pathology, the individual functioning as the technical supervisor must meet one of the following requirements: (m)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located and-- (m)(1)(ii) Be certified in anatomic pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (m)(2) Be certified in oral pathology by the American Board of Oral Pathology or possess qualifications for such certification; or (m)(3) An individual qualified under 493.1449(b) or paragraph (m)(1) or (2) of this section may delegate to an individual who is a resident in a training program leading to certification specified in paragraphs (b) or (m)(1) or (2) of this section, the responsibility for examination and interpretation of oral pathology specimens. (n) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of radiobioassay, the individual functioning as the technical supervisor must-- (n)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (n)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (n)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (n)(2)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing for the specialty of radiobioassay; or (n)(3)(i) Have an earned doctoral degree in a chemical, physical, biological or clinical laboratory science from an accredited institution; and (n)(3)(ii) Have at least 1 year of laboratory training or experience, or both, in high complexity testing within the specialty of radiobioassay; or (n)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (n)(4)(ii) Have at least 2 years of laboratory training or experience, or both, in high complexity testing for the specialty of radiobioassay; or (n)(5)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (n)(5)(ii) Have at least 4 years of laboratory training or experience, or both, in high complexity testing for the specialty of radiobioassay. (o) If the laboratory performs tests in the specialty of histocompatibility, the individual functioning as the technical supervisor must either-- (o)(1)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (o)(1)(ii) Have training or experience that meets one of the following requirements: (o)(1)(ii)(A) Have 4 years of laboratory training or experience, or both, within the specialty of histocompatibility; or (o)(1)(ii)(B)(1) Have 2 years of laboratory training or experience, or both, in the specialty of general immunology; and (o)(1)(ii)(B)(2) Have 2 years of laboratory training or experience, or both, in the specialty of histocompatibility; or (o)(2)(i) Have an earned doctoral

degree in a biological or clinical laboratory science from an accredited institution; and (o)(2)(ii) Have training or experience that meets one of the following requirements: (o)(2)(ii)(A) Have 4 years of laboratory training or experience, or both, within the specialty of histocompatibility; or (o)(2)(ii)(B)(1) Have 2 years of laboratory training or experience, or both, in the specialty of general immunology; and (o)(2)(ii)(B)(2) Have 2 years of laboratory training or experience, or both, in the specialty of histocompatibility. (p) If the laboratory performs tests in the specialty of clinical cytogenetics, the individual functioning as the technical supervisor must-- (p)(1)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (p)(1)(ii) Have 4 years of training or experience, or both, in genetics, 2 of which have been in clinical cytogenetics; or (p)(2)(i) Hold an earned doctoral degree in a biological science, including biochemistry, or clinical laboratory science from an accredited institution; and (p)(2)(ii) Have 4 years of training or experience, or both, in genetics, 2 of which have been in clinical cytogenetics. (q) If the requirements of paragraph (b) of this section are not met and the laboratory performs tests in the specialty of immunohematology, the individual functioning as the technical supervisor must-- (q)(1)(i) Be a doctor of medicine or a doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (q)(1)(ii) Be certified in clinical pathology by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (q)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (q)(2)(ii) Have at least one year of laboratory training or experience, or both, in high complexity testing for the specialty of immunohematology. Note: The technical supervisor requirements for "laboratory training or experience, or both" in each specialty or subspecialty may be acquired concurrently in more than one of the specialties or subspecialties of service. For example, an individual, who has a doctoral degree in chemistry and additionally has documentation of 1 year of laboratory experience working concurrently in high complexity testing in the specialties of microbiology and chemistry and 6 months of that work experience included high complexity testing in bacteriology, mycology, and mycobacteriology, would qualify as the technical supervisor for the specialty of chemistry and the subspecialties of bacteriology, mycology, and mycobacteriology.

This STANDARD is not met as evidenced by:

Based on review of personnel records and interview with the technical supervisor (TS) #2, the laboratory failed to ensure that 2 of 4 technical supervisors (TS) met the requirements to perform technical supervision and scientific oversight for the tests completed in the laboratory from 10/07/2020 to date of survey. Findings include: 1. On day of survey 07/11/2023 at 09:39 am, review of Personnel Competency records revealed TS #2 (CMS-209) reviewed competency assessments for personnel performing Histology/Cytology in June of 2022 and 2023. 2. Per "Delegation of Duties of the Laboratory Medical Director" document, the TS #2 is responsible for ensuring personnel have been appropriately trained and demonstrate competency prior to testing specimens. 3. Review of personnel credentials, revealed TS #2 does not meet the education requirements to assess competency in Histology/Cytology. TS#2 has a Bachelors of science in Medical Technology 4. Review of Supervisor Competency Record for TS #4 (CMS-209) completed in June of 2023 showed Manuals of Responsibility as Blood Bank, Chemistry and Hematology/Coagulation. 5. Per "Delegation of Duties of the Laboratory Medical Director" document, the

Laboratory Supervisor has permission to sign off on Laboratory outstanding specimen logs, manual entry logs, maintenance sheets, procedure manuals, worksheets and other paperwork otherwise signed by the Laboratory Medical Director. 6. Review of personnel credentials revealed TS #4 does not meet the education and experience requirements to supervise testing in the laboratory. 7. Interview with TS#2 confirmed the above findings on 07/12/2023 at 12:00 pm.