

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  39D0185430	<b>(X3) Date Survey Completed</b>  06/15/2020
<b>Name of Provider or Supplier</b>  Psh Cmg Andrewspatel West	<b>Street Address, City, State</b>  3912 Trindle Rd, Camp Hill, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5209</b>	<p><b>PERSONNEL COMPETENCY ASSESSMENT POLICIES</b> CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on laboratory personnel interview with the Laboratory Supervisor and Technical Supervisor-2 and review of the competency records, the Laboratory failed to follow written policies to assess employee competency for one of 2 Technical Supervisors, from (03/09/2018) through the date of the survey (06/15/2020). Findings: 1. Review of the competency records revealed no documentation of competency for Technical Supervisor-2 from (03/09/2018) through the date of the survey (06/15/2020). 2. Laboratory policies and procedures reviewed at the time of the survey (GenLAB6 version 1), revealed that the Technical Supervisor was required to be evaluated yearly by the Laboratory Director. 3. During the survey (09:54 06/12/2019), the Laboratory Supervisor confirmed, the above findings.</p>