

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0187192	(X3) Date Survey Completed 11/04/2022
Name of Provider or Supplier Dallastown Medical Associates	Street Address, City, State 1010 Blymire Rd, Dallastown, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5215	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on a review of the American Proficiency Institute (API) proficiency testing (PT) records and an interview with the Technical Consultant (TC), the laboratory failed to verify the accuracy for 2 of 6 Vaginal wet preparation, 1 of 6 Alanine transaminase (ALT), and 1 of 6 Methadone PT results in 2021 and 2022. Findings Include: 1. On the day of the survey, 11/04/2022 at 11:19 AM, a review of 2021 and 2022 API proficiency testing records revealed, the laboratory did not verify the accuracy of the following PT results for receiving a 'Not graded' score due to no consensus in 2021 and 2022: -2021 Hematology and Coagulation-Vaginal We Preparation KOH 3rd Event. -2022 Hematology and Coagulation-Vaginal We Preparation KOH 2nd Event. -2021 Chemistry Miscellaneous- Methadone 1st Event. -2021 Chemistry Core- ALT 1st Event. 2. TC #2 confirmed the findings above on 11/04/2022 around 12:15 PM.</p>