

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0187659	(X3) Date Survey Completed 02/02/2018
Name of Provider or Supplier Fertility Center Llc	Street Address, City, State 130 Leader Heights Road, York, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a February 01, 2018 review of CASPER 155 Report and performance evaluations from the proficiency testing organization, College of American Pathologist (CAP) the laboratory failed to successfully participate in a proficiency testing program approved by CMS for the analyst: HCG of the speciality Endocrinology. The laboratory had unsatisfactory scores for the 1st and 3rd event of 2017. See D2107</p>
D2107	<p>ENDOCRINOLOGY CFR(s): 493.843(f)</p>

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a February 01, 2018 review of CASPER 155 Report and performance evaluations from the proficiency testing organization, College of American Pathologist (CAP) the laboratory failed to successfully participate in a proficiency testing program approved by CMS for the analyst: hCG of the speciality Endocrinology. The laboratory had unsatisfactory scores for the 1st and 3rd event of 2017. Findings Include: 1. CAP 2017 Event 1 HCG score 20% 2. CAP 2017 Event 3 HCG score 0%