Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 39D0189070	**(X3) Date Survey Completed** 08/13/2025
Name of Provider or Supplier Bucktail Medical Center	**Street Address, City, State** 1001 Pine Street, South Renovo, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	A desk review for proficiency testing results was conducted by the Pennsylvania State Agency for Bucktail Medical Center on 08/13/2025. The laboratory was found out of compliance with the following conditions: 493.803 Condition: Successful participation.
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on review of the CASPER 0155 report and graded results from the proficiency testing (PT) organization, Wisconsin State Laboratory of Hygiene (WSLH), the

	laboratory failed to successfully participate in PT for Unexpected Antibody Identification (ABID). The laboratory had unsatisfactory scores for the 1st and 2nd Events of 2025. Refer to D2164 and D2172.
D2164	UNEXPECTED ANTIBODY DETECTION CFR(s): 493.861(a) (a) Failure to attain an overall testing event score of at least 100 percent is unsatisfactory performance. This STANDARD is not met as evidenced by: Based on review of the CASPER 155 report and graded results from the Wisconsin State Laboratory of Hygiene (WSLH) proficiency testing (PT) organization, the laboratory failed to attain an overall testing event score of at least 100 percent for 2 of 2 WSLH PT events in 2025 for Unexpected Antibody Identification (ABID) in the specialty of Immunohematology. The laboratory had unsatisfactory scores for the 1st and 2nd events of 2025. Findings Include: 1. Review of the CASPER 0155 report revealed the following unsatisfactory scores: - 2025 Event 1 ABID: 0% - 2025 Event 2 ABID: 80% 2. Further review of the laboratory's 2025 WSLH PT agency's graded results confirmed the above findings of failure to attain an overall testing score of at least 100 percent for 2 of 2 WSLH Events in 2025 ABID testing (Immunohematology).
D2172	UNEXPECTED ANTIBODY DETECTION CFR(s): 493.861(e) (e) Failure to achieve an overall testing event score of satisfactory for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on review of the CASPER 0155 Report and graded results from the proficiency testing (PT) organization, Wisconsin State Laboratory of Hygiene (WSLH), the laboratory failed to achieve an overall testing score of satisfactory performance for 2 of 2 WSLH PT events in 2025 for Unexpected Antibody Identification (ABID) testing (Immunohematology). The laboratory had unsatisfactory scores for the 1st and 2nd events of 2025. Findings include: 1. Review of the CASPER 0155 report revealed the following unsatisfactory scores: - 2025 Event 1 ABID: 0% - 2025 Event 2 ABID: 80% 2. Further review of the laboratory's 2025 WSLH PT agency's graded results confirmed the above findings resulting in unsuccessful performance for ABID testing (Immunohematology) for 2 of 2 WSLH Events in 2025.