Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 39D0189583	**(X3) Date Survey Completed** 01/15/2025
Name of Provider or Supplier Pottsville Cancer Clinic	**Street Address, City, State** 700 Schuylkill Manor Road Suite 7, Pottsville, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D2006	TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b) (b)The laboratory must examine or test, as applicable, the proficiency testing samples it receives from the proficiency testing program in the same manner as it tests patient specimens. This testing must be conducted in conformance with paragraph (b)(4) of this section. If the laboratory's patient specimen testing procedures would normally require reflex, distributive, or confirmatory testing at another laboratory, the laboratory should test the proficiency testing sample as it would a patient specimen up until the point it would refer a patient specimen to a second laboratory for any form of further testing. This STANDARD is not met as evidenced by: Based on review of the laboratory's American Proficiency Institute (API) proficiency testing (PT) records and interview with Testing Personnel (TP) # 1 (CMS 209 personnel # 2), the laboratory failed to examine 6 of 6 API PT hematology /coagulation testing events in the same manner as patient specimens in 2023 and 2024. Findings include: 1. On the day of survey, 01/15/2025 at 1:19 pm, review of the laboratory's API PT records revealed the laboratory failed to examine the following API hematology/coagulation PT samples in the same manner as patient specimens for 6 of 6 API PT testing events in 2023 and 2024: - Complete blood count (CBC) with 3-part white blood cell differential (HSY 01-HSY 05, HSY 06-HSY 10, HSY 11-HSY 15). 2. Interview with TP # 1 on 01/15/2025 at 1:30 pm confirmed the PT samples for CBC were analyzed by two different testing personnel (CMS 209 personnel #2 and #3) and the final result reported was based on the calculated average. 3. The laboratory could not provide documentation stating the CBC testing was included in the laboratory's patient repeat criteria. 4. TP # 1 confirmed the findings above on 01/15 /2025 at 2:31 pm.