Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0193262	(X3) Date Survey Completed 03/22/2023
Name of Provider or Supplier Medical Oncology Assocs Of Wyoming Vly	**Street Address, City, State** 382 Pierce Street, Kingston, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5775	COMPARISON OF TEST RESULTS CFR(s): 493.1281(a)(c) (a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites. (c) The laboratory must document all test result comparison activities. This STANDARD is not met as evidenced by: Based on lack of documentation and interview with testing personnel #1 (TP), the laboratory failed to evaluate twice a year the relationship between test results for 2 of 2 Cell-Dyn Emerald hematology analyzers from 01/14/2021 to the day of the survey. Findings include: 1. On the day of the survey, 03/22/2023 at 11:00 am, the laboratory could not provide documentation of the biannual comparison of test results between 2 of 2 Cell-Dyn Emerald hematology analyzers for the complete blood count (CBC) differentials performed from 01/14/2021 to 03/22/2023. 2. The laboratory performed 79,320 hematology tests in 2022 (annual volume listed on CMS-116). 3. TP #1 confirmed the finding above on 03/22/2023 at 2:00 pm.