

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0196241	(X3) Date Survey Completed 10/10/2023
Name of Provider or Supplier Alliance Cancer Specialists	Street Address, City, State 9910 E Roosevelt Blvd, Philadelphia, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5213	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(1)</p> <p>The laboratory must verify the accuracy of any analyte or subspecialty without analytes listed in subpart I of this part that is not evaluated or scored by a CMS-approved proficiency testing program.</p> <p>This STANDARD is not met as evidenced by: Based on review of the American Proficiency Institute's (API) proficiency testing (PT) records and interview with Testing Personnel (TP) #2 (CMS 209), the laboratory failed to verify the accuracy of analytes that were not evaluated or scored for 2 of 3 API hematology/coagulation PT events in 2023. Findings include: 1. On the day of survey, 10/10/2023 at 12:15 PM, review of the laboratory's API PT records revealed that the laboratory did not verify the accuracy for the following hematology /coagulation API events that were not graded in 2023: -API 1st Event 2023: Blood Cell ID- Blast, Neutrophil band, Dacrococyte, Neutrophil hyposegmented. -API 2nd Event 2023: Blood Cell ID- Lymph Reactive, Neutrophil segmented, Red blood cell normal, Platelet normal. 2. The TP #2 confirmed the findings above on 10/10/2023 at 16:15 PM</p>