

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0202793	(X3) Date Survey Completed 05/22/2018
Name of Provider or Supplier Advocare Main Line Pediatrics	Street Address, City, State 100 Deerfield Lane, Suite 110, Malvern, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory procedure manual (PM), throat culture (TC) log, and interview with the nurse manager, the laboratory failed to follow TC agar plate incubation procedures in May 2018. Findings include: 1. The laboratory TC PM (also the Collection, Culture, and Interpretation of a Throat Culture Screen for Presumptive Identification of Group A Streptococci, CDC) states "Re-incubate negative agar plates for an additional 24 hours and document final results." 2. The laboratory TC log (05/07/2018-05/11/2018) showed the laboratory streaked 22 specimens, from which a summary reads as follows: Read Result (hours) Streak Date Specimen 24 48 05/07/2018 4 Negative No Data 05/08/2018 6 Negative No Data 05/09/2018 5 Negative No Data 05/10/2018 4 Negative No Data 05/11/2018 3 Negative No Data 3. Total 22 TCs tested negative at 24 hours, but none (22 of 22) re-incubated and re-tested at 48 hours. 4. The laboratory reported all these 22 TC test results electronically. 5. The nurse manager confirmed above findings on 05/22/2018 at 10:00 AM.</p>