

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  39D0657582	<b>(X3) Date Survey Completed</b>  03/22/2023
<b>Name of Provider or Supplier</b>  Angeline Elizabeth Kirby Memorial Hlth Ctr	<b>Street Address, City, State</b>  71 North Franklin Street, Wilkes Barre, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5775</b>	<p>COMPARISON OF TEST RESULTS CFR(s): 493.1281(a)(c)</p> <p>(a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites. (c) The laboratory must document all test result comparison activities.</p> <p>This STANDARD is not met as evidenced by: Based on lack of documentation and interview with the technical supervisor (TS), the laboratory failed to evaluate twice a year the relationship between test results for 2 of 2 Atomic Absorption Spectrophotometers from 01/14/2021 to the day of the survey. Findings include: 1. On the day of the survey, 03/22/2023 at 11:00 am, the laboratory could not provide documentation of the biannual comparison of test results between 2 of 2 Perkin Elmer PinAAcle 900 Atomic Absorption Spectrophotometers for lead testing performed from 01/14/2021 to 03/22/2023. 2. The laboratory performed 14,760 lead examinations in 2022 (annual volume listed on CMS-116). 3. The TS confirmed the finding above on 03/22/2023 at 11:15 am.</p>