

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0674305	(X3) Date Survey Completed 07/18/2023
Name of Provider or Supplier Alliance Cancer Spec - Hem Onc	Street Address, City, State 1078 W Baltimore Pk, Suite 1, Media, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5221	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(d)</p> <p>All proficiency testing evaluation and verification activities must be documented.</p> <p>This STANDARD is not met as evidenced by: Based on review of the American Proficiency Institute (API) proficiency testing (PT) records and interview with the testing personnel(TP)#2, the laboratory failed to document the evaluation and verification activities for 1 of 3 events of hematology for the Unacceptable Proficiency Testing (PT) in 2022. Findings Include: 1. On the day of survey, 07/18/2023 at 10:20 am, a review of API proficiency testing records revealed that the laboratory received an unacceptable grade for: - 80% on API 2022 3rd event Hematology: Red cell count (Hem-3) (x10E12/L) Coulter AC-Tdiff II sample Hem-12. 2. The laboratory failed to provide documentation of the evaluation and verification activities for the above PT event. 3. The laboratory have an annual test volume of 26, 825 in hematology (CMS 116) 4. The TP#2 confirmed the findings above on 07/18/2023 at 11:15 am.</p>