

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0677979	(X3) Date Survey Completed 04/04/2022
Name of Provider or Supplier Clearfield Professional Group Ltd	Street Address, City, State 820 Turnpike Ave, Clearfield, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a review of the CASPER Report 155 report and graded results from the proficiency testing organization Medical Laboratory Evaluation (MLE), the laboratory failed to successfully participate in proficiency testing for the analyte Chloride, Sodium, Blood Urea Nitrogen, Hemoglobin, TSH and White Blood Cells, and specialties Endocrinology and Hematology. Refer to D2096, D2107, D2108, D2130 and D2131.</p>
D2096	<p>ROUTINE CHEMISTRY CFR(s): 493.841(f)</p>

	<p>Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the CASPER 155 report and graded results from the proficiency testing organization Medical Laboratory Evaluation (MLE), the laboratory failed to successfully participate in a proficiency testing for the Routine Chemistry analytes listed below. The laboratory had unsatisfactory scores for the 2nd event of 2021, 3rd event of 2021 and the 1st event of 2022. Findings include: Analyte Year Event Score% CHLORIDE 2021 3 40 CHLORIDE 2022 1 0 SODIUM 2021 3 20 SODIUM 2022 1 0 BUN 2021 2 60 BUN 2022 1 0</p>
D2107	<p>ENDOCRINOLOGY CFR(s): 493.843(f)</p> <p>Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the CASPER 155 report and graded results from the proficiency testing organization Medical Laboratory Evaluation (MLE), the laboratory failed to successfully participate in proficiency testing for the analyte:Thyroid Stimulating Hormone (TSH). The laboratory had unsatisfactory scores for the 3rd event of 2021 and the 1st event of 2022. Findings include: Analyte Year Event Score TSH 2021 3 20%. TSH 2022 1 0%.</p>
D2108	<p>ENDOCRINOLOGY CFR(s): 493.843(g)</p> <p>Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the CASPER Report 155 report and graded results from the proficiency testing organization Medical Laboratory Evaluation (MLE), the laboratory failed to achieve an overall testing score of satisfactory performance in the subspecialty Endocrinology. The laboratory had unsatisfactory scores for the 3rd event of 2021, and 1st event of 2022. Findings include: Subspecialty Year Event Score% Endocrinology 2021 3 60 Endocrinology 2022 1 0</p>
D2130	<p>HEMATOLOGY CFR(s): 493.851(f)</p> <p>Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.</p>

This STANDARD is not met as evidenced by:
Based on a review of the CASPER 155 report and graded results from the proficiency testing organization Medical Laboratory Evaluation (MLE), the laboratory failed to successfully participate in proficiency testing for the analyte: White Blood Cell (WBC), and Hemoglobin (HGB). The laboratory had unsatisfactory scores for the 2nd event of 2021, and 1st event 2022. Findings include: Analyte Year Event Score % HGB 2021 2 0 HGB 2022 1 0 WBC 2021 2 40 WBC 2022 1 0

D2131

HEMATOLOGY
CFR(s): 493.851(g)

Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:
Based on a review of the CASPER Report 155 report and graded results from the proficiency testing organization Medical Laboratory Evaluation (MLE), the laboratory failed to achieve an overall testing score of satisfactory performance in the Specialty Hematology. The laboratory had unsatisfactory scores for the 2nd event of 2021, and 3rd event of 2021. Findings include: Subspecialty Year Event Score Hematology 2021 2 73% Hematology 2022 1 0%