

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0683201	(X3) Date Survey Completed 02/03/2021
Name of Provider or Supplier Crumay Parnes Associates Inc	Street Address, City, State 1822 Good Hope Rd, Enola, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on laboratory personnel interview with Office Manager & the Laboratory Director and review of the personnel records, the Technical Consultant failed to establish written policies and procedures to assess Testing Personnel competency for 5 of 6 Testing Personnel. from 07/12/2018 through the date of the survey. Findings: 1. The laboratory could not produce written policies and procedures to assess Testing Personnel competency for KOH, Scabies, and Tzanck Smears, from 07/12/2018 through the date of the survey 02/03/2021. 2. The Laboratory Director acted as Technical Consultant#1 from 08/31/2018, through the date of the survey. 3. Review of the personnel records revealed no documentation of Testing Personnel competency for 5 of 6 Testing Personnel from 07/12/2018, through the date of the survey. 4. During the survey (11:30 02/03/2021), the Laboratory Supervisor confirmed, the above findings.</p>