

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0860538	(X3) Date Survey Completed 10/22/2018
Name of Provider or Supplier Bucks County Gastroenterology Assoc	Street Address, City, State 301 Oxford Valley Rd #701, Yardley, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6106	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(14)</p> <p>The laboratory director must ensure that an approved procedure manual is available to all personnel responsible for any aspect of the testing process.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory procedure manual and interview with the Laboratory Administrator at the time of the survey (10:00 10/22/18), the Laboratory Director failed to ensure an approved procedure manual was available for all aspects of the testing process, from 06/21/2017 through 10/22/2018. Findings include: 1. At the time of the survey (10:00 10/22/2018), review of the laboratory manual revealed that the laboratory failed to have a written policy for quality assessment. 2. During the survey, the Laboratory Administrator confirmed there was no written policy for quality assessment.</p>