

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0887030	(X3) Date Survey Completed 05/04/2021
Name of Provider or Supplier Yardley Dermatology Associates, Pc	Street Address, City, State 803 Floral Vale Boulevard, Yardley, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5449	<p>CONTROL PROCEDURES CFR(s): 493.1256(d)(3)(ii)(g)</p> <p>Unless CMS Approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub. 7), that provides equivalent quality testing, the laboratory must-- At least once a day patient specimens are assayed or examined perform the following for-- Each qualitative procedure, include a negative and positive control material; (g) The laboratory must document all control procedures performed.</p> <p>This STANDARD is not met as evidenced by: Based on review of quality control records and interview with testing personnel #12 (TP12), the laboratory failed to include a positive and a negative control materials each day of performing Potassium Hydroxide (KOH) microscopic examination on patient specimens for 2020 Findings include: 1. On the date of survey, 05/04/2021, the laboratory could not provide the documentation of quality control performed on the KOH microscopic examinations for 2020. 2. Two specimens were examined for KOH test in 2020. 3. The TP12 confirmed the findings above on 05/04/2021 at 10:50 a.m.</p>
D6168	<p>TESTING PERSONNEL CFR(s): 493.1487</p> <p>The laboratory has a sufficient number of individuals who meet the qualification requirements of 493.1489 of this subpart to perform the functions specified in 493.1495 of this subpart for the volume and complexity of testing performed.</p> <p>This CONDITION is not met as evidenced by: Based on review of the CLIA's Laboratory Personnel Report (Form CMS-209), review of personnel qualification records, and interview with the Testing Personnel</p>

(TP)#12, the laboratory failed to ensure that each individual performing High Complexity testing (2 of 13 TP) is qualified. Refer to D6171

D6171

TESTING PERSONNEL QUALIFICATIONS

CFR(s): 493.1489(b)

(b) Meet one of the following requirements: (b)(1) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located or have earned a doctoral, master's or bachelor's degree in a chemical, physical, biological or clinical laboratory science, or medical technology from an accredited institution; (b)(2)(i) Have earned an associate degree in a laboratory science, or medical laboratory technology from an accredited institution or-- (b)(2)(ii) Have education and training equivalent to that specified in paragraph (b)(2)(i) of this section that includes-- (b)(2)(ii)(A) At least 60 semester hours, or equivalent, from an accredited institution that, at a minimum, include either-- (b)(2)(ii)(A)(1) 24 semester hours of medical laboratory technology courses; or (b)(2)(ii)(A)(2) 24 semester hours of science courses that include-- (b)(2)(ii)(A)(2)(i) Six semester hours of chemistry; (b)(2)(ii)(A)(2)(ii) Six semester hours of biology; and (b)(2)(ii)(A)(2)(iii) Twelve semester hours of chemistry, biology, or medical laboratory technology in any combination; and (b)(2)(ii)(B) Have laboratory training that includes either of the following: (b)(2)(ii)(B)(1) Completion of a clinical laboratory training program approved or accredited by the ABHES, the CAHEA, or other organization approved by HHS. (This training may be included in the 60 semester hours listed in paragraph (b)(2)(ii)(A) of this section.) (b)(2)(ii)(B)(2) At least 3 months documented laboratory training in each specialty in which the individual performs high complexity testing. (b)(3) Have previously qualified or could have qualified as a technologist under 493.1491 on or before February 28, 1992; (b)(4) On or before April 24, 1995 be a high school graduate or equivalent and have either-- (b)(4)(i) Graduated from a medical laboratory or clinical laboratory training program approved or accredited by ABHES, CAHEA, or other organization approved by HHS; or (b)(4)(ii) Successfully completed an official U.S. military medical laboratory procedures training course of at least 50 weeks duration and have held the military enlisted occupational specialty of Medical Laboratory Specialist (Laboratory Technician); (b)(5)(i) Until September 1, 1997-- (b)(5)(i)(A) Have earned a high school diploma or equivalent; and (b)(5)(i)(B) Have documentation of training appropriate for the testing performed before analyzing patient specimens. Such training must ensure that the individual has-- (b)(5)(i)(B)(1) The skills required for proper specimen collection, including patient preparation, if applicable, labeling, handling, preservation or fixation, processing or preparation, transportation and storage of specimens; (b)(5)(i)(B)(2) The skills required for implementing all standard laboratory procedures; (b)(5)(i)(B)(3) The skills required for performing each test method and for proper instrument use; (b)(5)(i)(B)(4) The skills required for performing preventive maintenance, troubleshooting, and calibration procedures related to each test performed; (b)(5)(i)(B)(5) A working knowledge of reagent stability and storage; (b)(5)(i)(B)(6) The skills required to implement the quality control policies and procedures of the laboratory; (b)(5)(i)(B)(7) An awareness of the factors that influence test results; and (b)(5)(i)(B)(8) The skills required to assess and verify the validity of patient test results through the evaluation of quality control values before reporting patient test results; and (b)(5)(i)(B)(8)(ii) As of September 1, 1997, be qualified under 493.1489(b)(1), (b)(2), or (b)(4), except for those individuals qualified under paragraph (b)(5)(i) of this section who were performing high complexity testing on or before April 24, 1995; (b)(6) For blood gas analysis-- (b)(6)(i) Be qualified under 493.1489(b)(1), (b)(2), (b)(3), (b)(4), or (b)(5); (b)(6)(ii) Have

earned a bachelor's degree in respiratory therapy or cardiovascular technology from an accredited institution; or (b)(6)(iii) Have earned an associate degree related to pulmonary function from an accredited institution; or (b)(7) For histopathology, meet the qualifications of 493.1449 (b) or (l) to perform tissue examinations.

This STANDARD is not met as evidenced by:

Based on review of the CLIA's Laboratory Personnel Report (Form CMS-209), review of personnel qualification records, and interview with the Testing Personnel (TP)#12, the laboratory failed to ensure that each individual performing High Complexity testing (2 of 13 TP) had the minimum qualifications required from May, 2019 to the day of survey. Findings Include: 1. On the date of survey 05/04/2021 at 09:15 am, the surveyor reviewed the credentials of the testing personnel listed on the CMS 209 (TP12 and TP13) and discovered that 2 of 2 testing personnel who performed grossing for Mohs micrographic examinations from May, 2019 through the day of survey, did not have the minimum qualifications. 2. The TP12 confirmed the finding above on 05/04/2021 at 11:25 a.m. .