

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0894217	(X3) Date Survey Completed 11/02/2021
Name of Provider or Supplier Abington Surgical Center	Street Address, City, State 2701 Blair Mill Road#17, Willow Grove, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory procedure manual and interview with the Laboratory Director (LD), the laboratory failed to establish a competency assessment procedure to assess the competency of 5 of 6 Clinical Consultants (CC) from 11/2/2019 to 11/2/2021. Findings include: 1. On the day of survey 11/02/2021, the laboratory could not provide a written procedure to assess the competency of Clinical Consultants from 11/02/2019 to 11/02/2021. 2. The laboratory could not provide documentation of competency assessed for 5 of 6 CC. 3. The LD confirmed the findings above on 11/02/2021 at 9:40 a.m.</p>
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory procedure manual, review of peer review records, and interview with Laboratory Director (LD), the laboratory failed to include a peer review procedure for Hematoxylin and Eosin (H&E) microscopic examinations from</p>

11/02/2019 to the day of survey Findings Include: 1. On the day of survey, 11/02/2021, the LD could not provide a peer review procedure manual for H&E microscopic examinations from 11/02/2019 to the day of survey. 2. The LD confirmed the finding above on 11/02/2021 at 10:13 a.m.