

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0941130	(X3) Date Survey Completed 06/09/2023
Name of Provider or Supplier Pennsylvania Gastroenterology Consult	Street Address, City, State 899 Poplar Church Road, Camp Hill, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's peer review records, and interview with the histotechnologist (HT), the laboratory failed to ensure that the verification of accuracy for microscopic examinations for histopathology were performed at least twice annually, as required for tests not included in subpart I in 2022. Findings Include: 1. On the day of the survey, 06/09/2023 at 10:30 am, the laboratory could not provide documentation that the verification of accuracy for histopathology slide examinations stained using immunohistochemical stains, special stains, and hematoxylin and eosin (H&E) were performed at least twice annually in 2022. 2. The laboratory performed 11,506 histopathology examinations in 2022 (annual volume listed on form CMS-116). 3. The HT confirmed the findings above on 06/09/2023 around 12:00 pm.</p>