

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  39D0946050	<b>(X3) Date Survey Completed</b>  01/11/2022
<b>Name of Provider or Supplier</b>  Berkley Hills Lab	<b>Street Address, City, State</b>  322 Warren Street, Johnstown, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2094</b>	<p>ROUTINE CHEMISTRY CFR(s): 493.841(e)</p> <p>(1) For any unsatisfactory analyte or test performance or testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure. (2) For any unacceptable analyte or testing event score, remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two years from the date of participation in the proficiency testing event.</p> <p>This STANDARD is not met as evidenced by: Based on personnel interview of the Laboratory Supervisor and review of proficiency records, the laboratory failed to attain a satisfactory score for the following analytes in the American Academy of Family Physicians (AAFP) 3rd event of 2021. findings include: 1. PSA 50% 2. Vitamin B-12 50% 3. Urine Microalbumin 0% During the survey (16:30 01/11/2022), the Laboratory Supervisor confirmed the above findings.</p>