

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0975103	(X3) Date Survey Completed 03/12/2026
Name of Provider or Supplier Allegheny Reproductive Health Center	Street Address, City, State 5910 Kirkwood Street, Pittsburgh, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D3009	<p>FACILITIES CFR(s): 493.1101(c)</p> <p>The laboratory must be in compliance with applicable Federal, State, and local laboratory requirements.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's American Proficiency Institute (API) proficiency testing (PT) records and interview with the Laboratory Director (LD), the laboratory failed to maintain acceptable ranges statistically determined by the PT agency (API) for 2 of 3 events of Hematology/Coagulation PT performed in 2025. Findings include: 1. The PA State regulation 5.62(a) states, "Results of proficiency tests shall be maintained in the acceptable ranges statistically determined for each evaluation. Failure of a laboratory to satisfactorily perform in a proficiency test may be cause for revocation of approval of the specific tests involved." 2. On the day of survey, 03/12/2026 at 1:00 pm, review of API PT records revealed the laboratory scored 0% for urinalysis analytes in the following 2 of 3 events: - 2025: 2nd event (Glucose) - 2025: 3rd event (Glucose) 3. The LD confirmed the findings above on 03/12/2026 at 1:20 pm.</p>