

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D0978140	(X3) Date Survey Completed 08/22/2023
Name of Provider or Supplier Azizkhan Internal Medicine Associates	Street Address, City, State 888 Poplar Church Road, Camp Hill, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5213	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(1)</p> <p>The laboratory must verify the accuracy of any analyte or subspecialty without analytes listed in subpart I of this part that is not evaluated or scored by a CMS-approved proficiency testing program.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the American Proficiency Institute (API) proficiency testing (PT) records and interview with testing personnel (TP) #1 (CMS 209), the laboratory failed to verify the accuracy of the nongraded result obtained for 1 of 3 API PT events for Chemistry in 2022. Findings Include: 1. On the day of survey, 08/22/2023 at 12:17 PM, a review of the laboratory's API PT records revealed that the laboratory did not verify the accuracy for the following analyte in Chemistry that were not graded by the PT agency: -API 2022 Chemistry Core 1st Event-TIBC 2. The TP #1 confirmed the findings above on 08/22/2023 at 02:19 PM.</p>
D5221	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(d)</p> <p>All proficiency testing evaluation and verification activities must be documented.</p> <p>This STANDARD is not met as evidenced by: Based on review of the American Proficiency Institute (API) proficiency testing (PT) records and interview with the testing personnel (TP) #1 (CMS 209), the laboratory failed to document the evaluation and verification activities for 1 of 3 events of Chemistry for the Unacceptable Proficiency Testing (PT) performance in 2022. Findings Include: 1. On the day of survey, 08/22/2023 at 12:06 PM, a review of API proficiency testing records revealed that the laboratory received an unacceptable grade</p>

for: - 80% on API 2023 2nd event Chemistry Core: Glycated Hemoglobin. 2. The laboratory failed to provide documentation of the evaluation and verification activities for the above PT event. 3. TP #1 confirmed the findings above on 08/22/2023 at 02:19 PM.