

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  39D1005462	<b>(X3) Date Survey Completed</b>  07/18/2024
<b>Name of Provider or Supplier</b>  Alveoli Corporation DbA Lungs At Work	<b>Street Address, City, State</b>  5000 Waterdam Plaza Drive, Suite 180, McMurray, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2088</b>	<p>ROUTINE CHEMISTRY CFR(s): 493.841(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's American Proficiency Institute (API) proficiency testing (PT) records and interview with Testing Personnel (TP) #2, the laboratory failed to attain an overall testing event score of at least 80 percent for 1 of 3 chemistry testing events in 2023. Findings include: 1. On the day of survey 7/18/2024 at 12:30 pm, review of the laboratory's API PT records revealed the laboratory scored a zero percent (0%) for each of the following analytes in the 2nd event of 2023: - pCO<sub>2</sub>, pH, pO<sub>2</sub>, Carboxyhemoglobin, Hemoglobin, Methemoglobin and Oxyhemoglobin. 2. The laboratory failed to obtain an overall testing event score of at least 80 percent and resulted in unsatisfactory performance for the 2nd API PT event of blood gas analysis in 2023. 3. TP #2 confirmed the findings above on 7/18/2024 at 1:05 pm.</p>
<b>D2093</b>	<p>ROUTINE CHEMISTRY CFR(s): 493.841(d)</p> <p>Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's American Proficiency Institute (API) proficiency testing (PT) records and interview with Testing Personnel (TP) #2, the laboratory</p>

failed to return PT results to API within the timeframe specified for 1 of 3 chemistry testing events in 2023. Findings include: 1. On the day of survey 7/18/2024 at 12:30 pm, review of the laboratory's API PT records revealed the laboratory failed to participate and resulted in an unsatisfactory performance score of zero for the 2nd event of blood gas analysis in 2023. 2. Interview with TP #2 on 7/18/2024 at 1:05 pm confirmed the laboratory failed to return PT results to API within the specified timeframe for the 2nd event in 2023.