

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  39D1056834	<b>(X3) Date Survey Completed</b>  10/15/2020
<b>Name of Provider or Supplier</b>  Laboratory Corp Of America	<b>Street Address, City, State</b>  240 Middletown Boulevard, Suite 204, Langhorne, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5211</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based a review of proficiency testing records and interview with the Laboratory Supervisor on the date of the survey (10/15/2020), the Laboratory failed to ensure review and evaluation of the results obtained on proficiency testing for 1 of 2 American Proficiency Institute (API) 2020 events. . Findings: 1. The Director's review signature for proficiency test reports was not found for API chemistry proficiency event 2 in 2020. 2. The laboratory could not provide documentation of evaluation &amp; review performed on the 2020 API chemistry event 2. 3. During the survey (10:30 10 /15/2020), the Laboratory Supervisor confirmed the above findings.</p>