

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D2033974	(X3) Date Survey Completed 03/14/2018
Name of Provider or Supplier Greentree Medical Center	Street Address, City, State 995 Greentree Rd, Pittsburgh, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on laboratory procedure review and interview with the assigned laboratory testing person, laboratory testing personnel did not follow the laboratory's procedure manual for Urine Microscopic examination . Findings include: 1 At the time of the survey, (13:30 03/14/2018) the procedure for Urine Microscopic examination requires urine to be centrifuged between 2,000 and 3,000 revolutions per minute(rpm). 2 The only centrifuge in the lab was set above 3,000 rpm for serum separation, and can not be adjusted by the end user. 3 During the survey the assigned testing person confirmed the above findings.</p>