

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D2033974	(X3) Date Survey Completed 05/23/2024
Name of Provider or Supplier Greentree Medical Center	Street Address, City, State 995 Greentree Rd, Pittsburgh, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5215	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on review of the American Proficiency Institute (API) proficiency testing (PT) records and interview with testing personnel (TP) #1 (CMS 209), the laboratory failed to verify the accuracy of the PT results obtained for 1 of 3 API Microbiology testing events in 2023. Findings Include: 1. On the day of survey, 05/23/2024 at 12:02 pm., review of the laboratory's API PT records revealed that the laboratory did not verify the accuracy for the following analytes that were not scored by the PT agency in 2023: - 1st event: Influenza A (RSP - 04) 2. The API Proficiency Testing performance Evaluation form states " Laboratories are responsible for documenting and performing corrective action for failures and must perform a self-evaluation using statistics presented in the Participant Data Summary for samples that have not been graded." 3. TP #1 confirmed the findings above on 05/23/2024 at 2:30 pm.</p>