

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D2039545	(X3) Date Survey Completed 02/04/2020
Name of Provider or Supplier Pennsylvania Dermatology Partners, Inc	Street Address, City, State 2128 Penn Ave, 1st Floor, West Lawn, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on review of proficiency testing records and interview with the laboratory supervisor, the laboratory failed to verify at least twice annually accuracy (blind test) of Mohs microscopic examinations for 1 of 1 testing personnel (laboratory director) in 2018 and 2019. Findings Include: 1. The Proficiency Testing Policy states, "Dermatologists will circulate slides within the Pennsylvania Dermatology Partners, PC group and assess slides in a blind test". 2. On the day of survey, 02/04/2020, the laboratory could not provide documentation of blind testing conducted at least twice annually for 1 of 1 testing personnel who performed Mohs microscopic examinations in 2018 and 2019. 3. Approximately 1,977 Mohs microscopic examinations were analyzed in 2019. 4. The laboratory supervisor confirmed the findings above on 02/04/2020 around 8:45 am.