

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  39D2082546	<b>(X3) Date Survey Completed</b>  06/05/2018
<b>Name of Provider or Supplier</b>  Nj Cd DbA Dedicated Dermatology	<b>Street Address, City, State</b>  2895 Hamilton Blvd Suite 202, Allentown, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D6094</b>	<p><b>LABORATORY DIRECTOR RESPONSIBILITIES</b> CFR(s): 493.1445(e)(5)</p> <p>The laboratory director must ensure that the quality assessment programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur.</p> <p>This STANDARD is not met as evidenced by: Based on, the review of Laboratory Procedures and interview with Testing Personnel (TP) #2, the laboratory failed to ensure that quality assessment programs are maintained and documented to assure the quality of laboratory from 2017 to the date of survey. Findings: 1. On the day of survey, 06/05/2018, the laboratory provided a Quality Assurance Procedures, that was implemented in February of 2018. 2. The laboratory could not provide a Quality Assurance Procedure, signed by the Director before February 2018. From January 2017 to February 2018 (13 months) a Quality Assurance Procedures was not available on site. 3. TP#2 confirmed the findings above, on 06/05/2018 around 10:15 am.</p>