

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D2091155	(X3) Date Survey Completed 05/18/2023
Name of Provider or Supplier Dermatology Ltd, Chadds Ford	Street Address, City, State 1788 Wilmington West Chester Pike, Glen Mills, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6125	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(8)(v)</p> <p>The procedures for evaluation of the competency of the staff must include, but are not limited to assessment of test performance through testing previously analyzed specimens, internal blind testing samples or external proficiency testing samples.</p> <p>This STANDARD is not met as evidenced by: Based on review of Competency Assessment records and interview with Laboratory Director (LD), the Technical Supervisor (TS) failed to include assessment of test performance in the competency evaluation of 2 of 2 testing personnel who performed gross macroscopic examinations and specimen inking of frozen sections from 04/16 /2021 through date of survey. Findings include: 1. On the day of the survey 05/18 /2023 at 9:43am, review of Laboratory Personnel Competency Assessment Tracking form for Gross examinations and specimen inking of frozen sections, revealed the TS did not include assessment of test performance through previously analyzed specimens, internal blind testing samples or external proficiency testing samples from 04/16/2021 to 05/18/2023. 2. The laboratory reported an annual test volume of 2119 for Histopathology. (CMS-116) 3. LD confirmed the above findings on 05/18/2023 around 10:30am</p>