

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D2107253	(X3) Date Survey Completed 07/21/2022
Name of Provider or Supplier Pennsylvania Dermatology Partners-Pottstown	Street Address, City, State 2093 E High Street, Pottstown, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5429	<p>MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(a)(1)</p> <p>For unmodified manufacturer's equipment, instruments, or test systems, the laboratory must perform and document maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer.</p> <p>This STANDARD is not met as evidenced by: Based on observation of the laboratory and interview with the Laboratory Manager, the laboratory failed to establish a maintenance policy to assess the maintenance /function for 1 of 1 Fisher Scientific digital thermometer used to monitor the room temperature and humidity for Dermatopathology laboratory from January 9, 2019 to the day of survey. Findings Include: 1. On the day of survey, 07/21/2022 at 13:13 PM, the laboratory failed to provide a maintenance policy for 1 of 1 Fisher scientific thermometer (S/N 170015175). 2. An expiration date of 09/01/2019 was observed on the sticker posted on the Fisher Scientific digital thermometer. 3. Laboratory manager confirmed the findings above on 07/21/2022 at 13:25 PM.</p>