

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  39D2135089	<b>(X3) Date Survey Completed</b>  09/17/2025
<b>Name of Provider or Supplier</b>  Pshmg Camp Hill Dermatology	<b>Street Address, City, State</b>  425 N 21st Street, Camp Hill, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5401</b>	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>(a) A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on record review, lack of documentation, and interview with Technical Consultant (TC) #2, the laboratory failed to establish a written procedure that included the laboratory's process for performing and evaluating proficiency testing (PT) for Potassium Hydroxide (KOH) and Scabies microscopic examinations performed for 22 of 22 months from 10/17/2023 through the date of the survey. Findings include: 1. On the day of the survey, 09/17/2025 at 10:33 am, review of the laboratory's Wisconsin State Laboratory of Hygiene (WSLH) KOH PT records and procedure manual revealed the laboratory failed to establish a written procedure that included the laboratory's process for performing and evaluating PT testing for KOH and Scabies microscopic examinations performed from 10/17/2023 through 09/17/2025. 2. The laboratory reported an estimated annual test volume of 135 KOH Prep and Scabies examinations in 2024 (CMS 116, dated 09/11/2025). 3. TC#2 (CMS 209, dated 09/11/2025) confirmed the findings above on 09/17/2025 at 11:50 am.</p>