

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  39D2138528	<b>(X3) Date Survey Completed</b>  01/15/2026
<b>Name of Provider or Supplier</b>  Washington Health System Lab - MI	<b>Street Address, City, State</b>  80 Landings Drive 2nd Floor, Washington, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5413</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>(b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (b)(1) Water quality. (b)(2) Temperature. (b)(3) Humidity. (b)(4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory records, lack of documentation, and interview with the Laboratory Director (LD), the laboratory failed to monitor and document room temperature and humidity for 2 of 2 years to ensure reliable test system operation and test result reporting for instrumentation used to perform histopathology microscopic slide examinations from 01/09/2024 to the date of survey. Findings include: 1. The operating conditions listed in the instruction manual for the Olympus BX 46F microscope stated: "temperature: 5 C to 40C, humidity: 80%. RH max." 2. On the date of the survey, 01/15/2026 at 9:45 am, the laboratory could not provide documentation for monitoring room temperature and humidity to ensure operating conditions were met for 1 of 1 Olympus BX 46F microscope (s/n 7J40561) used to perform histopathology microscopic slide examinations for 2 of 2 years from 01/09/2024 to 01/15/2026. 3. Review of the laboratory's test logs revealed the laboratory performed 1 histopathology microscopic slide examination from 01/09/2024 to 01/15/2026. 4. The LD confirmed the findings above on 01/15/2026 at 11:05 am.</p>