Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D2142578	(X3) Date Survey Completed 06/08/2022
Name of Provider or Supplier Center For Emergency Medicine Of Western Pa, Inc	**Street Address, City, State** 10 Allegheny County Airport, West Mifflin, PA	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235 As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency. This STANDARD is not met as evidenced by: Based on review of Laboratory competency policy and interview with the Laboratory Director (LD), the laboratory failed to establish a procedure to assess the competency of 1 of 1 Clinical consultant (on the CMS 209 form, listed as personnel #2) for their responsibilities in 2020 and 2021. Findings include: 1. On the day of survey, 06/08 /2021 at 08:50 am, the LD, could not provide a policy that reviews how to assess the competency for 1 of 1 Clinical Consultant for their responsibilities in 2020 and 2021. 2. The laboratory could not provide documentation for competency assessment for 1 of 1 CC. 3. The LD confirmed the finding above on 06/08/2021 around 10:30 am.
D6051	TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(8)(v) The procedures for evaluation of the competency of the staff must include, but are not limited to assessment of test performance through testing previously analyzed specimens, internal blind testing samples or external proficiency testing samples. This STANDARD is not met as evidenced by: Based on review of competency assessment records, College of Pathology (CAP) proficiency testing (PT) records and interview with Laboratory Director (LD), the Technical Consultant (TC) failed to evaluate the test performance of 164 of 164 testing personnel (TP) through internal blind testing samples or external PT samples

	for blood gases assay examinations in 2020 and 2021. . Findings Include: 1. At the time of survey, 06/08/2022 at 09:30 am, the competency assessment records revealed the laboratory did not assess test performance of 164 of 164 TP through internal blind testing samples or external PT samples in 2020 and 2021 who performed blood gasses assay examinations on the ISTAT analyzer. 2. The LD confirmed the findings above on 06/08/2022 around 9:30 am.