

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D2142578	(X3) Date Survey Completed 03/06/2024
Name of Provider or Supplier Center For Emergency Medicine Of Western Pa, Inc	Street Address, City, State 10 Allegheny County Airport, West Mifflin, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5429	<p>MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(a)(1)</p> <p>For unmodified manufacturer's equipment, instruments, or test systems, the laboratory must perform and document maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer.</p> <p>This STANDARD is not met as evidenced by: Based on observation of the laboratory, lack of documentation, and interview with the laboratory director (LD) and compliance manager (CM), the laboratory failed to perform and document the maintenance/ function checks for 1 of 1 thermometer used to monitor refrigerator temperatures used for storage of chemistry reagents from 06/08/2022 to the day of survey. Findings Include: 1. On the day of survey, 03/06/2024 at 12:30 pm, the laboratory could not provide maintenance/function check records for the following 1 of 1 thermometer used to monitor refrigerator temperatures used for storage of chemistry reagents from 06/08/2022 to 03/06/2024: - Stanley S/N 000CCC1EE23F 2. The following Chemistry Blood Gas reagents were found to be stored in the laboratory. - 20 of 20 I stat CG4++ Cartridge (Manufacturer's storage requirements - 2 - 8 Celsius). 3. The LD and CM confirmed the findings above on 03/06/2024 at 1:45 pm.</p>