

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D2142589	(X3) Date Survey Completed 11/13/2025
Name of Provider or Supplier Aesthetic Dermatology Associates, Pc -Paoli Office	Street Address, City, State 4 Industrial Blvd, Suite 200, Paoli, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory procedures, lack of documentation and interview with Testing Personnel (TP) #3, the laboratory failed to verify twice annually the accuracy of microscopic examinations performed for mycology (potassium hydroxide) (KOH) and parasitology (scabies) for 2 of 2 years in 2024 and 2025. Findings include: 1. The laboratory's KOH/Scabies Examination procedure stated, "Physician shall maintain KOH quality assurance during yearly review. Surgeon has also set up review template with colleagues to review cases on an ongoing basis making sure it is done semiannually at least." 2. On the day of survey, 11/13/2025 at 12:14 pm, the laboratory could not provide documentation for the verification of accuracy performed at least twice annually for KOH and scabies microscopic examinations performed in 2024 and 2025. 3. Review of the laboratory's KOH/Scabies logs revealed the laboratory performed 2 KOH microscopic examinations in 2024 and 3 scabies microscopic examinations in 2025. 4. TP #3 (CMS 209 personnel #3, dated 11/03/2025) confirmed the findings above on 11/13/2025 at 1:47 pm.</p>