

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D2172816	(X3) Date Survey Completed 01/11/2022
Name of Provider or Supplier Schweiger Dermatology Group	Street Address, City, State 701 East Baltimore Pike, Ste #B, Kennett Square, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's procedures and interview with the General Manager (GM), the laboratory failed to have a policy to assess the competency of 2 of 3 Testing Personnel (TP) who performed potassium hydroxide (KOH) and Scabies microscopic analysis from 02/13/2020 to the date of survey. Findings include: 1. On the day of survey, 01/11/2022 at 09:27 a.m., the GM could not provide a competency assessment policy to assess the competency of 2 of 3 TP (CMS 209 personnel #2 and #3) from 02/13/2020 to the date of survey: 2. The laboratory could not provide competency assessment records for 2 of 3 TP who performed KOH and Scabies microscopic analysis in 2020 and 2021. 3. The GM confirmed the findings above on 01/11/2022 at 10:45 a.m.</p>