

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D2229796	(X3) Date Survey Completed 10/18/2023
Name of Provider or Supplier Schweiger Dermatology Group	Street Address, City, State 860 First Avenue, Suite 8b, King Of Prussia, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on lack of documentation, review of the laboratory procedure manual and interview with the General Manager (GM), the laboratory failed to perform twice annually the verification of accuracy of Mohs microscopic examinations from 12/15/2021 to the date of survey. Findings include: 1. On the day of survey, 10/18/2023 at 10:35 am, the laboratory could not provide the biannual documentation for the verification of accuracy of Mohs microscopic examinations from 12/15/2021 to the date of survey. 2. According to the laboratory's proficiency testing policy, two cases are to be submitted for peer review every 6 months. 3. The laboratory reported an annual volume of 265 microscopic examinations/tests performed in Histopathology (CMS 116). 4. The GM confirmed the above findings on 10/18/2023 at 10:55am.</p>