

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 39D2327596	(X3) Date Survey Completed 01/12/2026
Name of Provider or Supplier Dermatology Partners - Chambersburg	Street Address, City, State 1942 Scotland Ave, Chambersburg, PA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5781	CORRECTIVE ACTIONS CFR(s): 493.1282(b)(1) (b) The laboratory must document all corrective actions taken, including actions taken when any of the following occur: (b)(1) Test systems do not meet the laboratory's verified or established performance specifications, as determined in 493.1253(b), which include but are not limited to-- (b)(1)(i) Equipment or methodologies that perform outside of established operating parameters or performance specifications; (b)(1)(ii) Patient test values that are outside of the laboratory's reportable range of test results for the test system; and (b)(1)(iii) When the laboratory determines that the reference intervals (normal values) for a test procedure are inappropriate for the laboratory's patient population. This STANDARD is not met as evidenced by: Based on review of the laboratory's temperature logs, lack of documentation, and interview with the Director of Clinical Development (DCD), the laboratory failed to document corrective action taken when room temperatures were recorded above the laboratory's acceptable range of 68 to 76 degrees Fahrenheit for 5 of 23 days when microscopic dermatopathology examinations were performed from 08/25/2025 to 12/29/2025. Findings include: 1. On the day of the initial survey, 01/12/26 at 10:49 am, review of the laboratory's room temperature logs revealed room temperatures above 68 to 76 degrees Fahrenheit were recorded for the following days when microscopic dermatopathology examinations were performed from 08/25/2025 to 12/29/2025: - 11/10/25 (76.1 F) - 11/17/25 (76.5 F) - 12/1/25 (76.1 F) - 12/8/25 (76.1 F) - 12/15/25 (77.2 F) 2. The laboratory could not provide documentation of corrective actions taken when room temperatures exceeded the laboratory's established acceptable range. 3. Review of the laboratory's test logs revealed the laboratory performed 246 dermatopathology slide examinations from 08/25/2025 to 12/29/2025. 4. The DCD confirmed the above findings on 01/12/2026 at 12:10 pm.