Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 40D0985371	**(X3) Date Survey Completed** 08/05/2025
Name of Provider or Supplier Metro Pavia Clinic Bayamon	**Street Address, City, State** Santa Cruz St #77, Bayamon, PR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	The Centers for Medicare & Medicaid Services (CMS) conducted an unannounced CLIA Recertification survey at the Metro Pavia Clinic Bayamon on August 5, 2025. The laboratory was surveyed under 42 CFR part 493 CLIA Requirements. The following standard level deficiencies were found during the unannounced routine CLIA recertification survey ending on August 5, 2025.
D2098	ENDOCRINOLOGY CFR(s): 493.843(a) (a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event. This STANDARD is not met as evidenced by: Based on review of Puerto Rico Proficiency Testing (year 2025), CASPER Report 0155D scores, and interview with the laboratory supervisor on August 5, 2025 at 9:54 AM; it was determined that the laboratory failed to attain a score at least 80 percent in the triiodothyronine (T3 uptake) in the second proficiency testing event in the 2025 year. The finding includes: 1. The Puerto Rico Proficiency and Casper Report 0155D scores, showed that the laboratory obtained the following unsuccessful scores: Analyte: Sodium a. triiodothyronine (T3 uptake)- 0 % 2. The laboratory supervisor confirmed on August 5, 2025 at 9:54 AM; that the laboratory failed to attain a score at least 80 percent in the triiodothyronine (T3 uptake) in the second proficiency testing event in the 2025 year.
D6116	TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(3) (b)(3) Enrollment and participation in an HHS approved proficiency testing program commensurate with the services offered;

	This STANDARD is not met as evidenced by: Based on review of Puerto Rico Proficiency Testing scores (year 2025) and CASPER Report 0155D scores; it was determined that the technical supervisor failed to ensure that the laboratory achieved satisfactory participation in the second proficiency testing event of the year 2025. Refer to D2098.