

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  40D1095483	<b>(X3) Date Survey Completed</b>  04/30/2021
<b>Name of Provider or Supplier</b>  Laboratorio Clinico Riaza	<b>Street Address, City, State</b>  #502 Calle Riaza, Urb Matienzo Cintron, Bo Oriente, San Juan, PR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5014</b>	<p><b>GENERAL IMMUNOLOGY</b> CFR(s): 493.1208</p> <p>If the laboratory provides services in the subspecialty of General immunology, the laboratory must meet the requirements specified in 493.1230 through 493.1256, and 493.1281 through 493.1299.</p> <p>This CONDITION is not met as evidenced by: Based on Mycoplasma pneumoniae quality control records patient records review ( 2020-2021 ) and interview with the laboratory general supervisor on April 30, 2021 at 10:00 AM, it was determined that the laboratory failed to meet the requirements in the subspecialty of General Immunology. The finding includes: 1. The laboratory did not include each day of testing a negative and a positive control materials when patients serum specimens were tested for qualitative Mycoplasma IgM by the Immuno Card Mycoplasma method. Refer to : 5449- The laboratory did not include positive and negative control material</p>
<b>D5449</b>	<p><b>CONTROL PROCEDURES</b> CFR(s): 493.1256(d)(3)(ii)(g)</p> <p>Unless CMS Approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub. 7), that provides equivalent quality testing, the laboratory must-- At least once a day patient specimens are assayed or examined perform the following for-- Each qualitative procedure, include a negative and positive control material; (g) The laboratory must document all control procedures performed.</p> <p>This STANDARD is not met as evidenced by: Based on review of Mycoplasma pneumoniae IgM quality control results ( 2020-2021</p>

), patient test results records and interview with the laboratory supervisor on April 30, 2021 at 10:00 AM, it was found that the laboratory did not include a positive and a negative control material each day of patient testing. The findings include: 1. The laboratory began to perform patient's test for Mycoplasma pneumoniae on September 2020. 2. Review of the quality control and patient test results records showed that positive and negative controls were included when a new reagent box was opened. 3. The mycoplasma quality control records ( 2020-2021) and patient test results records showed that from February 12, 2020 to April 2021 the laboratory processed and reported 132 mycoplasma IgM . 2020-114 patient 2021 -18 patient 4. The laboratory supervisor stated on April 30, 2021 at 10:00A.M., that they included a negative and a positive control material when a new reagent box was opened and documented the procedural control with each patient.

**D6076**

**LABORATORY DIRECTOR**  
CFR(s): 493.1441

The laboratory must have a director who meets the qualification requirements of 493.1443 of this subpart and provides overall management and direction in accordance with 493.1445 of this subpart.

This CONDITION is not met as evidenced by:  
Based on general immunology quality control records review ( 2020-2021) and laboratory supervisor interview on April 30, 2021 at 10:00 A.M., it was determined that the laboratory director failed to fulfill his responsibilities and duties to ensure compliance with the laboratory quality control requirements. Refer to D6093

**D6093**

**LABORATORY DIRECTOR RESPONSIBILITIES**  
CFR(s): 493.1445(e)(5)

The laboratory director must ensure that the quality control programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur.

This STANDARD is not met as evidenced by:  
Based on general immunology quality control record review from February 2020 to April 2021 and laboratory general supervisor interview on April 30, 2021 at 10:00 A. M. it was determined that the laboratory director failed to ensure compliance with the requirement for analytic systems. The finding includes: 1. The laboratory did not include a positive and a negative control material when performed Mycoplasma IgM each day of patient testing. Refer to D5449.

**D6144**

**GENERAL SUPERVISOR RESPONSIBILITIES**  
CFR(s): 493.1463

The general supervisor is responsible for day-to-day supervision or oversight of the laboratory operation and personnel performing testing and reporting test results.

This STANDARD is not met as evidenced by:  
Based on general immunology quality control record review from February 2020 to April 2021 and laboratory general supervisor interview on April 30, 2021 at 10:00 A.

M. it was determined that the laboratory general supervisor failed to ensure compliance with the requirement for analytic systems. The finding includes: 1. The laboratory did not include a positive and a negative control material when performed Mycoplasma IgM each day of patient testing. Refer to D5449.