

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 40D2303458	(X3) Date Survey Completed 08/05/2026
Name of Provider or Supplier Lab Clinico Los Puertos - Magnolia	Street Address, City, State Urb Magnolia Gardens Calle 10 O-14, Bayamon, PR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The Centers for Medicare & Medicaid Services (CMS) conducted an unannounced CLIA Recertification survey at the Laboratorio Clinico Los Puertos-Magnolia on August 5, 2026. The laboratory was surveyed under 42 CFR part 493 CLIA Requirements. The following standard level deficiencies were found during the unannounced routine CLIA recertification survey ending on August 5, 2026.
D5215	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2) The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results). This STANDARD is not met as evidenced by: Based on review of Puerto Rico Proficiency Testing Service Program (PRPTSP) scores (years 2025 - 2026), Certification and Survey Provider Enhanced Reports (CASPER Report 0155D) scores, hematology Proficiency Testing (PT) scores (year 2025) and laboratory testing personnel interview on August 5, 2026, at 10:02 A.M.; the laboratory failed to evaluate the accuracy of testing in the hematology specialty when the laboratory received an artificially score of 100 percent from the PT provider. The laboratory processed and reported 331 patient samples from November 2025 through June, 2026. The findings include: 1. The CASPER Report 0155D scores were review on August 4, 2026 at 3:00 PM, and show that the laboratory received 100 percent score in the third event of hematology in the year 2025. 2. PRPTSP were reviewed from February 2025 through June 2026. 3. Review of the hematology PT scores for the third testing event in 2025 showed that the PT provider assigned an artificial score of 100 percent. The results were not evaluated. 4. During interview On August 5, 2026, at 10:02 A.M.; with the laboratory testing personnel, the accuracy of

	the excused hematology specialty (Complete Blood Count - (CBC) and White Blood Cell (WBC) 5 Parameters) was required. The laboratory testing personnel stated that no procedure for accuracy evaluation was performed. 5. The testing personnel also stated on August 5, 2026, at 10:02 A.M.; that no written procedure was developed by the laboratory to evaluated the accuracy of test not evaluated by the PT provider. 6. From November 2025 through June, 2026, the laboratory processed and reported 331 patient samples.
D6018	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(iii) (e)(4)(iii) All proficiency testing reports received are reviewed by the appropriate staff to evaluate the laboratorys performance and to identify any problems that require corrective action; and This STANDARD is not met as evidenced by: Based on review of Puerto Rico Proficiency Testing scores (year 2025-2026), CASPER Report 0155D scores, proficiency laboratory records and interview with the laboratory director on August 5, 2026 at 10:02 AM, the laboratory director failed to verify the accuracy of the hematology proficiency test when the lab obtained an artificial score. Refer to D5215.