

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 42D0249458	(X3) Date Survey Completed 10/04/2024
Name of Provider or Supplier Sandhills Pediatric & Adolescent Clinic	Street Address, City, State 1749 Marshall Street, Columbia, SC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a desk review of proficiency testing scores from the Casper Report 0155D and Wisconsin State Laboratory of Hygiene (WSLH) proficiency testing summary, the laboratory failed to successfully participate in a proficiency testing program approved by CMS. Findings included: 1. Review of Casper Report 0155D reveals failure to successfully participate in a CMS approved proficiency testing (PT) program by scores of 0% for 2 out of 3 events for 2023 as follows for White Blood Cell (WBC) Count: a. WSLH 2023 HemeReg1 score 0% b. WSLH 2023 HemeReg3</p>

score 0% 2. Review of CLIA regulations reveals that successful participation in a CMS approved PT program is testing scores of a minimum of 80% for each analyte tested.

D2123

HEMATOLOGY
CFR(s): 493.851(c)

Failure to participate in a testing event is unsatisfactory performance and results in a score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if-- (1) Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results; (2) The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and (3) The laboratory participated in the previous two proficiency testing events.

This STANDARD is not met as evidenced by:
Based on a desk review of proficiency testing scores from the Casper Report 0155D and Wisconsin State Laboratory of Hygiene (WSLH) proficiency testing summary, the laboratory unsuccessfully participated in a proficiency testing program subspecialty Hematology approved by CMS. Findings included: 1. Review of Casper Report 0155D reveals failure to successfully participate in a CMS approved proficiency testing (PT) program by scores of 0% for 2 of 3 events in 2023 as follows for White Blood Cell (WBC) Count: a. WSLH 2023 HemeReg1 score 0% b. WSLH 2023 HemeReg3 score 0% 2. Review of CLIA regulations reveals that successful participation in a CMS approved PT program is testing scores of a minimum of 80% for each analyte tested.

D2130

HEMATOLOGY
CFR(s): 493.851(f)

Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:
Based on a desk review of proficiency testing scores from the Casper Report 0155D and Wisconsin State Laboratory of Hygiene (WSLH) proficiency testing summary, the laboratory failed to successfully participate in a proficiency testing program subspecialty Hematology approved by CMS. Findings included: 1. Review of Casper Report 0155D reveals failure to successfully participate in a CMS approved proficiency testing (PT) program by scores of 0% for events 2 out of 3 events for 2023 for the following White Blood Cell (WBC) Count: a. WSLH 2023 HemeReg1 score 0% b. WSLH 2023 HemeReg3 score 0% 2. Review of CLIA regulations reveals that successful participation in a CMS approved PT program is testing scores of a minimum of 80% for each analyte tested.