

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 42D0252422	(X3) Date Survey Completed 10/10/2019
Name of Provider or Supplier Kendall & Kemmerlin Pa (Drs)	Street Address, City, State 1817 Woodruff Road, Greenville, SC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: During an onsite recertification survey on 10/10/2019, based on direct observation and testing personnel interview, it was determined that the laboratory failed to ensure that potassium hydroxide (KOH) solution was not used past the expiration date. Findings include: 1. During the laboratory walkthrough at 11:45 am, a bottle of KOH was observed near the microscope. The bottle was printed with a note to "discard after 2013". 2. During the exit interview at 12:00 pm, testing personnel confirmed that the KOH on site had expired and was actively being utilized for patient testing..</p>