Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	42D0891427	09/25/2023
Name of Provider or Supplier	**Street Address, City, State**	
Charleston Pain And Rehabilitation Center	1124 Sam Rittenberg Boulevard Suite 1, Charleston, SC	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Recertification Survey was initiated on 09/25/2023 and concluded on 09/25/2023. The facility was found not to be in compliance with the laboratory requirements of 42 CFR Part 493 with deficiencies cited.
D6121	TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(8)(i) The procedures for evaluation of the competency of the staff must include, but are not limited to direct observations of routine patient test performance, including patient preparation, if applicable, specimen handling, processing and testing. This STANDARD is not met as evidenced by: Based on interview and document and policy review, the laboratory failed to ensure an annual competency evaluation was documented in 2022 for 1 (Testing Personnel [TP] #1) of 1 testing personnel reviewed. Findings included: A review of a facility policy titled "Personnel Competency Assessment" signed 04/15/2015 by the Laboratory Director revealed, "The competence, qualifications, and responsibilities of all laboratory personnel will be assessed on a bi-annual and annual basis." Further review revealed, "Evaluation will be performed at 6 months after the techs [sic] start date and 1 year after the start date. Evaluation will be performed annually from then on. Each laboratory testing personnel will have a file maintained in the lab with the following information:" which included "Semi-Annual and Annual Personnel Evaluation." A review of an annual competency evaluation form for TP #1 revealed that the last documented competency evaluation was dated 12/20/2021. During an interview on 09/25/2023 at 11:20 AM, TP #1 stated their competency evaluation was performed annually in December. Regarding a competency evaluation being performed in December 2022, TP #1 stated the Laboratory Director (LD) assessed her competency by observing as she performed laboratory testing, but did not document or date on an evaluation form.