

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  42D0919829	<b>(X3) Date Survey Completed</b>  08/23/2021
<b>Name of Provider or Supplier</b>  Colonial Family Practice	<b>Street Address, City, State</b>  325 Broad Street, Sumter, SC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2016</b>	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: During a proficiency testing desk review performed on 08/20/2021, based on review of CASPER report 155D and graded reports from American Acadamey of Family Physicians (AAFP), it was determined that the laboratory failed to successfully participate in proficiency testing for the specialty of chemistry, the analyte albumin and chloride for three consecutive proficiency testing events reviewed (2020, Events 1,2 and 3) (See D2087, D2096).</p>
<b>D2087</b>	<p>ROUTINE CHEMISTRY CFR(s): 493.841(a)</p>

Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.

This STANDARD is not met as evidenced by:

During the proficiency testing desk review performed on 08/20//2021, based on review of CASPER report 155D and graded proficiency reports from AAFP, it was determined that the laboratory failed to attain a satisfactory score of at least 80% for albumin and chloride on three consecutive proficiency testing events reviewed (2020, Events 1, 2, and 3) . Findings include: 1. The CASPER 155D report revealed the following scores for your laboratory's albumin: a. 2020, Event 1: 40% b. 2020, Event 2: 0% c. 2020, Event 3: 0% The CASPER 155D report revealed the following scores for your laboratory's chloride: a. 2020, Event 1: 0% b. 2020, Event 2: 0% c. 2020, Event 3: 0% 2. The scores were confirmed by review of the graded AAFP results. Scores less than 80% for these analytes indicate failure or unsatisfactory performance. A failure of the analytes for two consecutive or two out of three testing events is scored as unsuccessful. A failure of the analyte for three consecutive or three out of four/five events is scored as a repeat unsuccessful.

**D2096**

**ROUTINE CHEMISTRY**  
CFR(s): 493.841(f)

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

During a proficiency testing desk review performed on 08/20/2021, based on review of the CASPER 155D report and graded report from AAFP, it was determined that the laboratory failed to achieve satisfactory performance for albumin and chloride in three consecutive testing events (2020, Events 1,2 and 3). See D2087.

**D6000**

**MODERATE COMPLEXITY LABORATORY DIRECTOR**  
CFR(s): 493.1403

The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.

This CONDITION is not met as evidenced by:

During the proficiency testing desk review performed on 08/20/2021, based on review of CASPER report 155D and graded AAFP reports, it was determined that the laboratory director failed to ensure proficiency testing for albumin, chloride was performed as required by 42 CFR, Part 493.801(see D6016).

**D6016**

**LABORATORY DIRECTOR RESPONSIBILITIES**  
CFR(s): 493.1407(e)(4)(i)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform

test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;

This STANDARD is not met as evidenced by:

During the desk review on 08/20/2021, based on CASPER report 155D review and AAFP graded report review, it was determined that the laboratory director failed to ensure the laboratory attained a result of 80% for albumin, chloride, and ALT, specialty of routine chemistry. The laboratory failed three consecutive proficiency testing events for albumin and chloride (2020, Events 1,2 and 3).