

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 42D1048788	(X3) Date Survey Completed 08/24/2018
Name of Provider or Supplier Rapha Family Footcare, Pa	Street Address, City, State 819 Carn Street, Walterboro, SC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: During an onsite recertification on 08/22/2018, based on lack of documentation and testing personnel interview, the laboratory failed to verify the accuracy of Dermatophyte Test Medium (DTM) culture testing at least twice annually for 1 of 2 years reviewed (2017). Findings include: 1. Documentation of a semi-annual test accuracy verification for the DTM culture testing for 2017 was unavailable for review on the the day of the survey . 2. Testing personnel confirmed during an onsite interview on 08/22/2018 at 2:00pm that the laboratory had failed to verify accuracy of DTM culture testing at least twice annually.</p>