

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 42D1060453	(X3) Date Survey Completed 01/30/2018
Name of Provider or Supplier Premier Medical Center, Inc	Street Address, City, State 5390 Dorchester Road, Charleston, SC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5777	<p>COMPARISON OF TEST RESULTS CFR(s): 493.1281(b)(c)</p> <p>(b) The laboratory must have a system to identify and assess patient test results that appear inconsistent with the following relevant criteria, when available: (b)(1) Patient age. (b)(2) Sex. (b)(3) Diagnosis or pertinent clinical data. (b)(4) Distribution of patient test results. (b)(5) Relationship with other test parameters. (c) The laboratory must document all test result comparison activities.</p> <p>This STANDARD is not met as evidenced by: During an onsite initial survey on 01/30/2018, based on the laboratory procedure manual, lack of documentation, and testing personnel interview, the laboratory failed to document all test result comparison activities for 8 of 12 months reviewed (May 2017 through December 2017). Findings include: 1. The laboratory procedure manual stated that 5 charts would be reviewed quarterly to confirm the correct relationship of patient information to patient test results. 2. There were no chart reviews available for review on the day of the survey for the months May 2017 through December 2017. 3. Testing personnel confirmed during an onsite interview on 01/30/2018 at 1:30pm that the laboratory had failed to document quarterly chart reviews as required.</p>