

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 42D1099166	(X3) Date Survey Completed 02/04/2025
Name of Provider or Supplier Carolinas Physicians Network Inc DbA Atrium Health	Street Address, City, State 7666 Charlotte Hwy, Ste 120, Indian Land, SC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An onsite CLIA recertification survey was conducted on February 4, 2025, at the clinical laboratory of Atrium Health Primary Care Barnett Family Medicine Carolina Lakes by the South Carolina Department of Public Health's (SC DPH) Bureau of Nursing Homes and Medical Services. The laboratory was found to be out of compliance with 42 CFR Part 493, CLIA Requirements for Laboratories. The following is a list of deficiencies as a result of CLIA survey:
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) (d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: BBBb