

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 42D2171061	(X3) Date Survey Completed 04/05/2021
Name of Provider or Supplier Ovp Health	Street Address, City, State 251 West Palmetto Street, Florence, SC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: During a PT Desk Review 03/11/2021, based on review of CASPER report 155D and graded reports from American Proficiency Institute (API) proficiency, it was determined that the laboratory failed to successfully participate in proficiency testing for the specialty of routine chemistry, the analyte Aspartate Aminotransferase (AST) for two consecutive proficiency testing events reviewed (2019 Event 3 and 2020 Event 1). See D2087, and D2096.</p>
D2087	<p>ROUTINE CHEMISTRY CFR(s): 493.841(a)</p>

Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.

This STANDARD is not met as evidenced by:

During the proficiency testing desk review performed on 05/30/2019, based on review of CASPER report 155D and graded proficiency reports from API, it was determined that the laboratory failed to attain a satisfactory score of at least 80% for Aspartate Aminotransferase (AST) on two consecutive proficiency testing events. Findings include: 1. The CASPER 155D report revealed the following scores for your laboratory's AST: a. 2019, Event 3: 0% b. 2020, Event 1: 0% 2. The scores were confirmed by review of the graded API results. Scores less than 80% for these analytes indicate failure or unsatisfactory performance. A failure of the analytes for two consecutive or two out of three testing events is scored as unsuccessful. A failure of the analyte for three consecutive or three out of four/five events is scored as a repeat unsuccessful.

D2096

ROUTINE CHEMISTRY

CFR(s): 493.841(f)

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

During a proficiency testing desk review performed on 03/11/2021, based on review of the CASPER 155D report and graded report from API, it was determined that the laboratory failed to achieve satisfactory performance for Aspartate Aminotransferase (AST) in two consecutive testing events (2019, Event 3 and 2020, Event 1). See D2087.