

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 42D2213939	(X3) Date Survey Completed 06/16/2026
Name of Provider or Supplier Clearline Dermatology	Street Address, City, State 776 Daniel Ellis Dr Ste 1a, Charleston, SC	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced onsite CLIA recertification survey was conducted on June 16, 2026, at the laboratory of Clearline Dermatology of Charleston by the South Carolina Department of Public Health (SC DPH) Bureau of Nursing Homes and Medical Services. The laboratory was found to be out of compliance with Medicare condition 42 CFR Part 493, CLIA requirements for laboratories. The following is the singular citation for the June 16, 2026 recertification survey:
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on records review, lack of documentation, and staff interview, the laboratory to at least twice annually verify Mohs reporting accuracy by using a comparison studies. Findings included: 1. Review of comparison study documentation reveals one (1) date annually for 3 out of 3 years reviewed (2024,2025, and 2026) 2. Review of the policy and procedure for "Dermatology Practice Laboratory Systems Quality Assessment (QA)" reveals a lack of requirement for QA frequency. 3. In an interview on June 16, 2026 at 12:30pm with the practice manager in the laboratory office, the findings were confirmed.